

From: Adam Molnar/ESC/R3/USEPA/US
Sent: 3/1/2012 4:20:00 PM
To: Eric Graybill/ESC/R3/USEPA/US@EPA
CC:
Subject: Fw: Ethylene Glycol technical help

Adam Molnar, Chemist
Environmental Science Center
701 Mapes Road
Fort Meade, MD 20755
molnar.adam@epa.gov
410-305-2676

----- Forwarded by Adam Molnar/ESC/R3/USEPA/US on 03/01/2012 04:19 PM -----

From: Wayne Whipple/R5/USEPA/US
To: Stevie Wilding/ESC/R3/USEPA/US
Cc: R3 ESC-LTIG ORGANIC
Date: 01/11/2012 10:50 AM
Subject: Re: Ethylene Glycol technical help

We had modified analysis for them at superfund CLP using direct injection on a carbowax column. We were using mass spec but an FID should work also. Not sure how the detection limits compare but either way good luck with lower ppm range with that one. Troy will probably tell you about his propylene glycol work.

Wayne J. Whipple, Ph. D.
Chemist
US EPA R5 Chicago Regional Laboratory
536 S Clark Street
10th Floor
Chicago, IL 60605
phone 312 353 9063
fax 312 582 5168

From: Stevie Wilding/ESC/R3/USEPA/US
To: R3 ESC-LTIG ORGANIC
Date: 01/11/2012 07:54 AM
Subject: Ethylene Glycol technical help

Region 5 or 6 -- could you let Adam know how you are analyzing ethylene glycol.

----- Forwarded by Stevie Wilding/ESC/R3/USEPA/US on 01/11/2012 08:52 AM -----

From: Adam Molnar/ESC/R3/USEPA/US
To: Stevie Wilding/ESC/R3/USEPA/US
Cc: Cynthia Caporale/ESC/R3/USEPA/US@EPA
Date: 01/11/2012 08:33 AM
Subject: Region 6 help with Ethylene Glycol

Stevie,

It says in the Dimock summary that you or Cindy will contact R5 and R6 today. I would like to know if they have a working ethylene glycol analysis by GC-FID. If they are getting good results I'd like to have a contact to discuss what they are doing.

Are you planning a conference call today?

Adam Molnar, Chemist
Environmental Science Center
701 Mapes Road
Fort Meade, MD 20755
molnar.adam@epa.gov
410-305-2676